

Batch Certificate For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL			
NAME OF PRODUCT	Plasma 1X (human-tech) DNA-free		
DESCRIPTION	Human serum proteins in common plasma concentrations (1-fold concentrated), electrolytes, EDTA		
CATALOG NUMBER	SID-000007		
BATCH NUMBER	00080		
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> • Manufactured und sealed in class 2 safety cabinet • Bottled with qualified Dispenser • At room temperature • Produced according to internal quality standards related to EN ISO 13485 		
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none"> • 2D barcoded tube with screw cap • Material: Polypropylen (PP) 		
DATE OF MANUFACTURE	22. July 2020		
EXPIRY DATE	21. July 2022		
CONCENTRATION	1X		
NOMINAL VOLUME	5,0 ml		
QUALITY	Sterile, DNA-free		
STORAGE CONDITIONS	+ 2-8 °C		
MANUFACTURING AND QUALITY CONTROL SITES	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality Control	Test Method	Acceptance criteria
	Quality (DNA-free) ¹ LOD	Real-Time PCR Investigator Quantiplex Pro Kit (Qiagen)	< 0,5 pg/μl ¹

RESULTS OF ANALYSIS		Result	PASS/FAIL
	Quality	< 0,5 pg/μl	PASS
COMMENTS/REMARKS	Not applicable		

Name and position/title of Person authorising the batch release:

Mr. Björn Nowack

Date of batch release:

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.