



Release Date 10.12.2020
Version / Index 3 1
Print Date 17.01.2022

BATCH CERTIFICATE

For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	Plasma 1X (human-tech) DNA-free		
DESCRIPTION	Human serum proteins in common plasma concentrations (1-fold concentrated), electrolytes, EDTA		
CATALOG NUMBER	SID-000007		
BATCH NUMBER	00272		
MANUFACTURING CONDITIONS	<ul style="list-style-type: none">· Manufactured und sealed in class 2 safety cabinet· Bottled with qualified Dispenser· At room temperature· Manufactured according to DIN EN ISO 13485:2016		
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none">· 2D barcoded tube with screw cap· Material: Polypropylen (PP)		
DATE OF MANUFACTURE	13.01.2022		
EXPIRY DATE	12.01.2024		
CONCENTRATION	1X		
NOMINAL VOLUME	5.0 ml		
QUALITY	DNA-free		
STORAGE CONDITIONS	+ 2-8 °C		
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria
	Quality (DNA-free) LOD	Real-Time PCR Investigator Quantiplex Pro Kit (Qiagen)	< 0.5 pg/µl
RESULTS OF ANALYSIS		Result	PASS / FAIL
	Quality	< 0.5 pg/µl	PASS
COMMENTS / REMARKS	not applicable		

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release: 17.01.2022

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.