

BATCH CERTIFICATE

For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	Plasma 1X (human-tech) DNA-free		
DESCRIPTION	Human serum proteins in common plasma concentrations (1-fold concentrated), electrolytes, EDTA.		
CATALOG NUMBER	SID-000007		
BATCH NUMBER	00617		
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> · Manufactured and sealed in class 2 safety cabinet · Bottled with qualified Dispenser · Manufactured according to DIN EN ISO 13485:2016 		
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none"> · 2D barcoded tube with screw cap · Material: Polypropylen (PP) 		
DATE OF MANUFACTURE	04.12.2023		
EXPIRY DATE	03.12.2025		
CONCENTRATION	1X		
NOMINAL VOLUME	5.0 ml		
QUALITY	DNA-free		
STORAGE CONDITIONS	+ 2-8°C		
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria
	Quality (DNA-free) LOD	Real-Time PCR Investigator Quantiplex Pro Kit (Qiagen)	< 0.5 pg/μl
RESULTS OF ANALYSIS		Result	PASS / FAIL
	Quality	< 0.5 pg/μl	PASS
COMMENTS / REMARKS	not applicable		

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release: 06.12.2023

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.