

## Batch Certificate For Research Use Only

### PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	Plasma 1x (human-tech) DNAfree		
DESCRIPTION	Human serum proteins in common plasma concentrations, electrolytes, EDTA		
CATALOG NUMBER	SID-0007		
BATCH NUMBER	00005		
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> <li>• Manufactured und sealed in class 2 safety cabinet</li> <li>• Bottled with qualified Dispenser</li> <li>• At room temperature</li> </ul>		
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none"> <li>• 2D barcoded tube with screw cap</li> <li>• Material: Polypropylen (PP)</li> </ul>		
DATE OF MANUFACTURE	12 Dec 2018		
EXPIRY DATE	Dec 2019		
CONCENTRATION	1x		
NOMINAL VOLUME	5,0 ml		
QUALITY	Sterile, DNAfree		
STORAGE CONDITIONS	+ 2-8 °C		
MANUFACTURING AND QUALITY CONTROL SITES	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality Control	Test Method	Acceptance criteria
	Quality (DNAfree)	Real-Time PCR Investigator Quantiplex Pro Kit (Qiagen)	< 0,5 pg/μl <sup>1</sup>
	<sup>1</sup> LOD		

RESULTS OF ANALYSIS	Result	PASS/FAIL
	Quality	< 0,5 pg/ $\mu$ l
COMMENTS/REMARKS	Not applicable	

Name and position/title of Person authorising the batch release:

Mr. Björn Nowack

Date of batch release: 18 Dec. 2018

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.