

## Batch Certificate For Research Use Only

### PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	5-Gene-Multiplex 0.1% AF cfDNA AKT1/BRAF/ERBB2/KRAS/PIK3CA in Plasma
DESCRIPTION	5-Gene-Multiplex 0.1% AF cfDNA AKT1/BRAF/ERBB2/KRAS/PIK3CA in highly characterized human DNA from cell lines. Human proteins in common plasma concentrations, electrolytes, EDTA, cfDNA / ctDNA in common plasma concentrations.
CATALOG NUMBER	SID-000088
BATCH NUMBER	00129
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> <li>• Manufactured and sealed in class 2 safety cabinet</li> <li>• Bottled with qualified liquid handling workstation</li> <li>• At room temperature</li> </ul>
PACKAGE SIZE	<ul style="list-style-type: none"> <li>• 2D barcoded tube with screw cap</li> </ul>
PACKAGE TYPE	<ul style="list-style-type: none"> <li>• Material: Polypropylen (PP)</li> </ul>
DATE OF MANUFACTURE	11.11.2020
EXPIRY DATE	10.11.2021
TARGET CONCENTRATION	80 ng/ml (dsDNA)
TARGET QUANTITY	400 ng (dsDNA)
NOMINAL VOLUME	5 ml
MUTATION	AKT1 p.E17K (COSM33765*, COSV62571334*, substitution, c.49G>A, Exon 2) BRAF p.V600E (COSM476*, COSV56056643*, substitution, c.1799T>A, Exon 15) ERBB2 p.E770_A771insAYVM (new: p.Y772_A775dup) (COSM20959*, COSV54062409*, insertion, c.2313_2324dup, Exon 19) KRAS p.G12D (COSM521*, COSV55497369*, substitution, c.35G>A, Exon 1) KRAS p.Q61K (COSM549*, COSV55502066*, substitution, c.181C>A, Exon 2) KRAS p.A146T (COSM19404*, COSV55501778*, substitution, c.436G>A, Exon 3) PIK3CA p.H1047R (COSM775*, COSV55873195*, substitution, c.3140A>G, Exon 20) PIK3CA p.E545K (COSM763*, COSV55873239* substitution, c.1633G>A, Exon 9) <small>* GRCh38 COSMIC v91</small>
ALLELE FREQUENCY	0.1%
QUALITY	DNA quantity metrologically traceable to internationally certified reference material <sup>1</sup>

<sup>1</sup> ERM\_AD442K  
Phone: +49 (0) 381 377 182 01

	The copy number values are metrologically traceable to the natural units count 1 and ratio 1 and International System of Units (SI) derived units of volume.																												
STORAGE CONDITIONS	+ 2 - 8 °C																												
MANUFACTURING AND QUALITY CONTROL SITES	SensID GmbH Schillingallee 68, 18057 Rostock, Germany																												
TEST METHOD AND ACCEPTANCE CRITERIA	Quality Control	Test Method	Acceptance Criteria																										
	Fragmentation	Fragment Length Analysis <sup>2</sup> Agilent High Sensitivity DNA Kit (Agilent Technologies)	peak size 167 bp ± 10% (151 bp – 181 bp)																										
	Quantification	dsDNA measurement: Qubit dsDNA BR Assay Kit (Invitrogen) dsDNA amount per ml plasma	80 ng/ml ± 10% (72-88 ng/ml)																										
	Allele Frequency	ddPCR Analysis using BioRad QX200™ System	AF 0.1% ±60% (0.04–0.16%)																										
RESULTS OF ANALYSIS		Result	PASS/FAIL																										
	Fragmentation	176 bp	PASS																										
	Quantity	80 ng/ml plasma	PASS																										
	Allele Frequency	<table border="1"> <thead> <tr> <th>Mutation</th> <th>AF in %</th> <th>PASS/FAIL</th> </tr> </thead> <tbody> <tr> <td>AKT1 E17K</td> <td>0.10</td> <td>PASS</td> </tr> <tr> <td>BRAF V600E</td> <td>0.11</td> <td>PASS</td> </tr> <tr> <td>ERBB2 E770_A771insAYVM (new: Y772_A775dup)</td> <td>0.13</td> <td>PASS</td> </tr> <tr> <td>KRAS G12D</td> <td>0.08</td> <td>PASS</td> </tr> <tr> <td>KRAS Q61K</td> <td>0.07</td> <td>PASS</td> </tr> <tr> <td>KRAS A146T</td> <td>0.09</td> <td>PASS</td> </tr> <tr> <td>PIK3CA H1047R</td> <td>0.14</td> <td>PASS</td> </tr> <tr> <td>PIK3CA E545K</td> <td>0.16</td> <td>PASS</td> </tr> </tbody> </table>	Mutation	AF in %	PASS/FAIL	AKT1 E17K	0.10	PASS	BRAF V600E	0.11	PASS	ERBB2 E770_A771insAYVM (new: Y772_A775dup)	0.13	PASS	KRAS G12D	0.08	PASS	KRAS Q61K	0.07	PASS	KRAS A146T	0.09	PASS	PIK3CA H1047R	0.14	PASS	PIK3CA E545K	0.16	PASS
Mutation	AF in %	PASS/FAIL																											
AKT1 E17K	0.10	PASS																											
BRAF V600E	0.11	PASS																											
ERBB2 E770_A771insAYVM (new: Y772_A775dup)	0.13	PASS																											
KRAS G12D	0.08	PASS																											
KRAS Q61K	0.07	PASS																											
KRAS A146T	0.09	PASS																											
PIK3CA H1047R	0.14	PASS																											
PIK3CA E545K	0.16	PASS																											

<sup>2</sup> Measured before filling in product tube  
Phone: +49 (0) 381 377 182 01

COMMENTS/REMARKS

ADDITIONAL INFORMATION:

**Copy numbers (CN) of the respective measurements**

Mutation	CN wt <sup>3</sup> /ml	CN mut <sup>4</sup> /ml
AKT1 E17K	8030	8
BRAF V600E	7173	8
ERBB2 E770_A771insAYVM (new: Y772_A775dup)	11853	16
KRAS G12D	9880	8
KRAS Q61K	11894	8
KRAS A146T	13600	12
PIK3CA H1047R	15190	21
PIK3CA E545K	10139	16

*Table 1 indicates the values of the QC assays performed by SensID GmbH with a DNA input of ~100 ng. The value for the respective mutation results from the mean value of three measured replicates (CN values are rounded). CN concentration values per milliliter (ml) plasma are based on droplet digital (ddPCR) assay counts dilution factors, and droplet volume measurements. The detection of the amount of CNs may vary depending on the assay used. Therefore, due to assay properties, there may be deviations in the observed number of copies and allele frequencies compared to the values given here.*

Name and position/title of Person authorising the batch release:

Mr. Björn Nowack, Managing Director

Date of batch release: 16.12.2020

Signature batch release: Björn Nowack

This document was created electronically and is valid without a signature.

<sup>3</sup> Wild Type

<sup>4</sup> Mutation