

Batch Certificate For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	5-Gene-Multiplex 0.1% AF cfDNA AKT1/BRAF/ERBB2/KRAS/PIK3CA
DESCRIPTION	5-Gene-Multiplex 0.1% AF cfDNA AKT1/BRAF/ERBB2/KRAS/PIK3CA is highly characterized human DNA from cell lines.
CATALOG NUMBER	SID-000092
BATCH NUMBER	00095
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> • Manufactured and sealed in class 2 safety cabinet • At room temperature
PACKAGE SIZE	<ul style="list-style-type: none"> • 2D barcoded tube with screw cap
PACKAGE TYPE	<ul style="list-style-type: none"> • Material: Polypropylen (PP)
DATE OF MANUFACTURE	01.09.2020
EXPIRY DATE	31.08.2022
CONCENTRATION	20 ng/μl (dsDNA)
QUANTITY	400 ng (dsDNA)
NOMINAL VOLUME	25 μl; (548 ng)
MUTATION	AKT1 p.E17K (COSM33765*, COSV62571334*, substitution, c.49G>A, Exon 2) BRAF p.V600E (COSM476*, COSV56056643*, substitution, c.1799T>A, Exon 15) ERBB2 p.E770_A771insAYVM (new: p.Y772_A775dup) (COSM20959*, COSV54062409*, insertion, c.2313_2324dup, Exon 19) KRAS p.G12D (COSM521*, COSV55497369*, substitution, c.35G>A, Exon 1) KRAS p.Q61K (COSM549*, COSV55502066*, substitution, c.181C>A, Exon 2) KRAS p.A146T (COSM19404*, COSV55501778*, substitution, c.436G>A, Exon 3) PIK3CA p.H1047R (COSM775*, COSV55873195*, substitution, c.3140A>G, Exon 20) PIK3CA p.E545K (COSM763*, COSV55873239* substitution, c.1633G>A, Exon 9) <small>* GRCh38 COSMIC v91</small>
ALLELE FREQUENCY	0.1%
QUALITY	DNA quantity metrologically traceable to internationally certified reference material ¹ The copy number values are metrologically traceable to the natural units count 1 and ratio 1 and International System of Units (SI) derived units of volume.
STORAGE CONDITIONS	+ 2 - 8 °C
MANUFACTURING AND QUALITY CONTROL SITES	SensID GmbH Schillingallee 68, 18057 Rostock, Germany

¹ ERM_AD442K
Phone: +49 (0) 381 377 182 01

TEST METHOD AND ACCEPTANCE CRITERIA	Quality Control	Test Method	Acceptance criteria
	Fragmentation	Fragment Length Analysis Agilent High Sensitivity DNA Kit (Agilent Technologies)	peak size 167 bp ± 10% (151 bp – 181 bp)
	Quantification	Total DNA measurement: Spectrophotometry ssDNA [ng/μl] = (A260-A320)*38 ^{2,3} dsDNA measurement: Qubit dsDNA BR Assay Kit (Invitrogen)	Total DNA: n.a. ⁴ dsDNA: 17.5 – 22.5 ng/μl
	Allele Frequency	ddPCR Analysis using BioRad QX200™ System	AF 0.1% ±60% (0.04-0.16%)

RESULTS OF ANALYSIS																													
	Result	PASS/FAIL																											
Fragmentation	174 bp	PASS																											
Quantity	28.3 ng/μl (total DNA) 21.9 ng/μl (dsDNA)	PASS																											
Allele Frequency	<table border="1"> <thead> <tr> <th>Mutation</th> <th>AF in %</th> <th>PASS/FAIL</th> </tr> </thead> <tbody> <tr> <td>AKT1 E17K</td> <td>0.10</td> <td>PASS</td> </tr> <tr> <td>BRAF V600E</td> <td>0.11</td> <td>PASS</td> </tr> <tr> <td>PIK3CA H1047R</td> <td>0.08</td> <td>PASS</td> </tr> <tr> <td>PIK3CA E545K</td> <td>0.08</td> <td>PASS</td> </tr> <tr> <td>ERBB2 E770_A771insAYVM (new: Y772_A775dup)</td> <td>0.10</td> <td>PASS</td> </tr> <tr> <td>KRAS G12D</td> <td>0.08</td> <td>PASS</td> </tr> <tr> <td>KRAS Q61K</td> <td>0.07</td> <td>PASS</td> </tr> <tr> <td>KRAS A146T</td> <td>0.09</td> <td>PASS</td> </tr> </tbody> </table>		Mutation	AF in %	PASS/FAIL	AKT1 E17K	0.10	PASS	BRAF V600E	0.11	PASS	PIK3CA H1047R	0.08	PASS	PIK3CA E545K	0.08	PASS	ERBB2 E770_A771insAYVM (new: Y772_A775dup)	0.10	PASS	KRAS G12D	0.08	PASS	KRAS Q61K	0.07	PASS	KRAS A146T	0.09	PASS
Mutation	AF in %	PASS/FAIL																											
AKT1 E17K	0.10	PASS																											
BRAF V600E	0.11	PASS																											
PIK3CA H1047R	0.08	PASS																											
PIK3CA E545K	0.08	PASS																											
ERBB2 E770_A771insAYVM (new: Y772_A775dup)	0.10	PASS																											
KRAS G12D	0.08	PASS																											
KRAS Q61K	0.07	PASS																											
KRAS A146T	0.09	PASS																											

²Protocol NK603 – Community Reference Laboratory for GM Food and Feed

³Measured before filling in product tube

⁴not applicable

COMMENTS/REMARKS

ADDITIONAL INFORMATION:

Copy numbers (CN) of the respective measurements

Mutation	CN wt ⁵ /μl	CN mut ⁶ /μl
AKT1 E17K	1956	2
BRAF V600E	1775	2
PIK3CA H1047R	3787	3
PIK3CA E545K	2398	2
ERBB2 E770_A771insAYVM (new: Y772_A775dup)	2885	3
KRAS G12D	2451	2
KRAS Q61K	2987	2
KRAS A146T	3341	3

Table 1 indicates the values of the QC assays performed by SensID GmbH with a DNA input of ~150 ng. The value for the respective mutation results from the mean value of three measured repetitions of QC Pool of 10 QC samples (CN values are rounded). CN concentration values per microliter (μl), are based on droplet digital (ddPCR) assay counts dilution factors, and droplet volume measurements. The detection of the amount of CNs may vary depending on the assay used. Therefore, due to assay properties, there may be deviations in the observed number of copies and allele frequencies compared to the values given here.

Name and position/title of Person authorising the batch release:

Mr. Björn Nowack, Managing Director

Date of batch release: 04.09.2020

Signature batch release: Björn Nowack

This document was created electronically and is valid without a signature.

⁵ Wild Type

⁶ Mutation