



Release Date Version / Index Print Date 12.02.2021 3 1 01.11.2021

## BATCH CERTIFICATE

For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL						
NAME OF PRODUCT	Ashkenazim Son	Ashkenazim Son FFPE Reference Standard				
DESCRIPTION	Human FFPE Reference Standard (curl)					
CATALOG NUMBER	SID-000100	SID-000100				
BATCH NUMBER	00254					
MANUFACTURING CONDITIONS	<ul> <li>Manufactured according to DIN EN ISO 13485:2016</li> <li>At room temperature</li> </ul>					
PACKAGE SIZE AND TYPE	<ul> <li>2D barcoded tube with screw cap</li> <li>Material: Polypropylen (PP)</li> </ul>					
DATE OF MANUFACTURE	05.10.2021					
EXPIRY DATE	04.10.2023					
FORMAT	10 µm section / curl					
MUTATION * GRCh38 COSMIC v91	PIK3CA p.E542K (COSV55873227*, substitution, c.1624G>A, Exon 9) PIK3CA p.H1047R (COSV55873195*, substitution, c.3140A>G, Exon 20)					
ALLELE FREQUENCY	0%					
QUALITY	DNA quantity metrologically traceable to internationally certified reference material (ERM_AD442K). The copy number values are metrologically traceable to the natural units count 1 and ratio 1 and International System of Units (SI) derived units of volume.					
STORAGE CONDITIONS	Short term storage (up to 30 days): 2-8°C Long term storage: -18 to -25°C					
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany					
TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria			
	Cell density	Visual	> 60%			
	Quantification	dsDNA measurement: Qubit** dsDNA BR Assay Kit (Invitrogen) RNA measurement: Qubit** RNA BR Assay Kit (Invitrogen)	dsDNA: > 400 ng RNA > 400 ng			
	Allele frequency	Allele frequency analysis by ddPCR** (BioRad QX200™)	AF 0.0%			
**Measured after extraction with Qiagen AllPrep DNA/RNA FFPE Kit						





RESULTS OF ANALYSIS	Quality control	Result		PASS / FAIL	
	Cell density	> 60%		PASS	
		563.2 ng (dsDNA)		PASS	
	Quantification	912.9 ng (RNA)			
	Allele frequency	Mutation	AF in %	PASS / FAIL	
		PIK3CA p.E542K PIK3CA p.H1047R	0.00 0.10	PASS PASS	
	Additional information:				
COMMENTS / REMARKS	1) Theoretical DNA yield from 1 curl (assumption diploid chromosome set) 2509 ng				
	2) Measurement of copy number				
MEASUREMENT OF COPY NUMBER	Mutation		CN wt/ng tracted DN	CN mut/ng A extracted DNA	
	PIK3CA p.E542K		482	0.5	
	PIK3CA p.H1047R		506	0	
	wt: wildtype; mut: mutation				
	The table above indicates the values of the QC assays performed by SensID GmbH with a DNA input of ~10 ng. The value for the respective mutation results from the mean value of QC samples according to ISO 2859-1:2014-08 (CN values are rounded). CN concentration values per nanogram extracted DNA are based on droplet digital (ddPCR) assay counts dilution factors, and droplet volume measurements. The detection of the amount of CNs may vary depending on the assay used. Therefore, due to assay properties, there may be deviations in the observed number of copies and allele frequencies compared to the values given here.				

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release:

26.10.2021

Björn Nowack

Signature batch release:

This document has been created electronically and is valid without signature.