



Release Date Version / Index Print Date 15.09.2022 3 1 16.11.2023

BATCH CERTIFICATE

For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL

	<u></u>	9=					
NAME OF PRODUCT	Ashkenazim Son FFPE Reference Standard						
DESCRIPTION	Human FFPE Reference Standard (curl)						
CATALOG NUMBER	SID-000100						
BATCH NUMBER	00599						
MANUFACTURING CONDITIONS	· Manufactured according to DIN EN ISO 13485:2016						
PACKAGE SIZE AND TYPE	2D barcoded tube with screw cap Material: Polypropylen (PP)						
DATE OF MANUFACTURE	02.11.2023						
EXPIRY DATE	01.11.2025						
FORMAT	10 μm section / curl						
MUTATION * GRCh38 COSMIC v97	PIK3CA p.E542K (COSV55873227*, substitution, c.1624G>A, Exon 9)						
	PIK3CA p.E545G (COSV55873220*, substitution, c.1634A>G, Exon 9)						
	PIK3CA p.H1047R (COSV55873195*, substitution, c.3140A>G, Exon 20)						
ALLELE FREQUENCY	0%						
QUALITY	DNA quantity metrologically traceable to internationally certified reference material (ERM_AD442K). The copy number values are metrologically traceable to the natural units count 1 and ratio 1 and International System of Units (SI) derived units of volume.						
STORAGE CONDITIONS	Short term storage (up to 30 days): + 2-8°C Long term storage: - 18 to - 25°C						
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany						
TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria				
	Cell density	Visual	> 60%				
	Quantification	dsDNA measurement: Qubit** dsDNA BR Assay Kit (Invitrogen)	dsDNA: > 400 ng				
	Quantification	RNA measurement: Qubit** RNA BR Assay Kit (Invitrogen)	RNA > 400 ng				
	Allele frequency	Allele frequency analysis by ddPCR** (BioRad QX200™)	AF: 0.00% (0.00 – 0.03%)				
	**Measured after extraction with Qiagen AllPrep DNA/RNA FFPE Kit						





Release Date 15.09.2022 Version / Index 3 1 Print Date 16.11.2023

RESULTS OF ANALYSIS	Quality control	Result			PASS / FAIL		
	Cell density	> 60%			PASS		
	-	868.6 ng (dsDNA)		()	DAGG		
	Quantification	1467.7 ng (RNA)			PASS		
		Mutation		AF in %	PASS / FAIL		
	Allala fraguanay	PIK3CA p.E542K		0.00	PASS		
	Allele frequency	PIK3CA p.E545G		0.00	PASS		
		PIK3CA p.H1047R	4 p.H1047R		PASS		
COMMENTS / REMARKS	1304 ng	1) Theoretical DNA yield from 1 curl (assumption diploid chromosome set) 1304 ng 2) Measurement of copy number					
MEASUREMENT OF COPY NUMBER	Mutation	Mutation		l wt/ng icted DNA	CN mut/ng extracted DNA		
	PIK3CA p.E542K	PIK3CA p.E542K		+83.04	0.00		
	PIK3CA p.E545G	PIK3CA p.E545G		426.23	0.00		
	PIK3CA p.H1047F	PIK3CA p.H1047R		¥57.80	0.00		
	wt: wildtype; mut: mutation						
		The table above indicates the values of the QC assays performed by SensID GmbH with a					
	,	DNA input of 1 ng. The value for the respective mutation results from the mean value of QC					
	,	samples according to ISO 2859-1:2014-08 (CN values are rounded). CN concentration values					
	, -	per nanogram extracted DNA are based on droplet digital (ddPCR) assay counts dilution					
	' '	factors, and droplet volume measurements. The detection of the amount of CNs may vary					
	, ,	depending on the assay used. Therefore, due to assay properties, there may be deviations in					
	the observed number	the observed number of copies and allele frequencies compared to the values given here.					

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release: 16.11.2023

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.