

Batch Certificate

For Research Use Only

PRODUCT	INFORMATION.	and Qualit'	Y CONTROL

NAME OF PRODUCT PIK3CA-E545G 50%AF FFPE Reference Standard

DESCRIPTION Human FFPE Reference Standard (curl)

CATALOG NUMBER | SID-000106

BATCH NUMBER 00074

MANUFACTURING

• Manufactured and sealed according to internal quality

CONDITIONS standards related to EN ISO 13485

At room temperature

PACKAGE SIZE AND

• 2D barcoded tube with screw cap

TYPE • Material: Polypropylen (PP)

DATE OF 01.04.2020

MANUFACTURE

EXPIRY DATE 31.03.2022

FORMAT 10 µm section / 1 curl

MUTATION PIK3CA p.E545G (COSM764*, COSV55873220*, substitution, c.1634A>G, Exon 9)

* GRCh38 COSMIC v91

ALLELIC FREQUENCY 50.0%

QUALITY DNA quantity metrologically traceable to internationally certified

reference material¹

The copy number values are metrologically traceable to the

natural units count 1 and ration 1 and International System of Units

(SI) derived units of volume.

STORAGE + 2-8 °C

CONDITIONS

MANUFACTURING AND | SensID GmbH

QUALITY CONTROL Schillingallee 68, 18057 Rostock, Germany

Schillingaliee 60, 10037 Nostock, Centrain

SITES

AF in %

50.6

PASS



Page 2/3

TEST METHOD AND	Quality Control	Test Method	Acceptance	
ACCEPTANCE			criteria	
CRITERIA	Cell Density	Visual	> 60 %	
CITILITIA		Agarose gel electrophoresis ²	Bright band of high-	
	Quality	1% Gel with fluorescent DNA	molecular-weight	
		stain in 1 % TAE buffer	gDNA≥20kb	
	-	dsDNA measurement ² : Qubit	dsDNA:	
	Quantification	dsDNA BR Assay Kit (Invitrogen)	> 400 ng	
		RNA measurement²: Qubit	RNA:	
		RNA BR Assay Kit (Invitrogen)	> 400 ng	
	Allelie Feerman	ddPCR Analysis²	AF 50.0 %	
	Allelic Frequency	using BioRad QX200™ System	(45.0-55.0 %)	
RESULTS OF ANALYSIS				
	Result		PASS/FAIL	
	Cell Density	Visual: > 60 %	PASS	
	Quality	Bright band of high-molecular-w	eight PASS	
	Quality	gDNA ≥ 20 kb	PASS	
	524.0 ng (dsDNA)			
	Quantity	446.9 ng (RNA)	PASS	
		ı ~		

COMMENTS/REMARKS

Additional information:

Theoretical DNA yield from 1 curl under the assumption of a diploid chromosome set:

3,024 ng (dsDNA)

Allelic

Frequency

Copy numbers (CN) of the respective measurements

Mutation

PIK3CA E545G

Table 1 indicates the values of the QC assays performed by SensID GmbH with a DNA input of ~10 ng. The value for the respective mutation results from the mean value of five measured replicates (CN values are rounded). CN values per nanogram extracted DNA, are based on droplet digital (ddPCR) assay counts dilution factors, and droplet volume measurements. The detection of the amount of CNs may vary depending on the assay used. Therefore, due to assay properties, there may be deviations in the observed number of copies and allele frequencies compared to the values given here.

Mutation	CN wt ³ / ng extracted DNA	CN mut ⁴ / ng extracted DNA
PIK3CA E545G	221	226

⁴ Mutation

Phone: +49 (0) 381 377 182 01

 $^{^2\,\}mathrm{Measured}$ after extraction with Qiagen AllPrep DNA/RNA FFPE Kit

³ Wild Type



Name and position/title of Person authorising the batch release:

Mr. Björn Nowack, Managing Director

Date of batch release: 03.07.2020

Signature batch release: Björn Nowack

This document was created electronically and is valid without a signature.

Phone: +49 (0) 381 377 182 01