

Batch Certificate For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL			
NAME OF PRODUCT	Plasma 2X (human-tech) DNA-free		
DESCRIPTION	Human serum proteins in common plasma concentrations (2-fold concentrated), electrolytes, EDTA		
CATALOG NUMBER	SID-000001		
BATCH NUMBER	00071		
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> • Manufactured und sealed in class 2 safety cabinet • Bottled with qualified Dispenser • At room temperature • Produced according to internal quality standards related to EN ISO 13485 		
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none"> • 2D barcoded tube with screw cap • Material: Polypropylen (PP) 		
DATE OF MANUFACTURE	09 June 2020		
EXPIRY DATE	08 June 2022		
CONCENTRATION	2X		
NOMINAL VOLUME	2,5 ml		
QUALITY	DNA-free		
STORAGE CONDITIONS	+ 2-8 °C		
MANUFACTURING SITES	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality Control	Test Method	Acceptance criteria
	Quality (DNA-free)	Real-Time PCR Investigator Quantiplex Pro Kit (Qiagen)	< 0,5 pg/μl ¹

¹ Limit of Detection

RESULTS OF ANALYSIS	Quality	Result	PASS/FAIL
		< 0,5 pg/ μ l	PASS
COMMENTS/REMARKS	-		

Name and position/title of Person authorising the batch release:

Mr. Björn Nowack

Date of batch release: 15 June 2020

Signature batch release: Björn Nowack

This document is produced electronically and is valid without signature.