

## BATCH CERTIFICATE

For Research Use Only

### PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	Plasma 2X (human-tech) DNA-free		
DESCRIPTION	Human serum proteins in common plasma concentrations (2-fold concentrated), electrolytes, EDTA		
CATALOG NUMBER	SID-000001		
BATCH NUMBER	00158		
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> <li>· Manufactured und sealed in class 2 safety cabinet</li> <li>· Bottled with qualified Dispenser</li> <li>· At room temperature</li> <li>· Manufactured according to DIN EN ISO 13485:2016</li> </ul>		
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none"> <li>· 2D barcoded tube with screw cap</li> <li>· Material: Polypropylen (PP)</li> </ul>		
DATE OF MANUFACTURE	20.01.2021		
EXPIRY DATE	19.01.2023		
CONCENTRATION	2X		
NOMINAL VOLUME	2.5 ml		
QUALITY	DNA-free		
STORAGE CONDITIONS	+ 2-8 °C		
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria <sup>1</sup>
	Quality (DNA-free) <sup>1</sup> LOD	Real-Time PCR Investigator Quantiplex Pro Kit (Qiagen)	< 0.5 pg/μl
RESULTS OF ANALYSIS		Result	PASS / FAIL
	Quality	< 0.5 pg/μl	PASS
COMMENTS / REMARKS	not applicable		

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release: 29.01.2021

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.