

Batch Certificate

For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT

cfDNA (human) AF: 0% Ashkenazim Son

DESCRIPTION

cfDNA (human) AF: 0% Ashkenazim Son is highly characterized

human DNA from cell lines.

CATALOG NUMBER

SID-000003

BATCH NUMBER

00063

MANUFACTURING

Manufactured and sealed in class 2 safety cabinet

CONDITIONS

At room temperature

PACKAGE SIZE AND

2D barcoded tube with screw cap

TYPF

Material: Polypropylen (PP)

DATE OF MANUFACTURE 04.05.2020

EXPIRY DATE

03.05.2022

CONCENTRATION

20 ng/µl (dsDNA)

QUANTITY

400 ng (dsDNA)

NOMINAL VOLUME

25 µl; (480 ng)

MUTATION

AKT1 p.E17K (COSM33765*, COSV62571334*, substitution, c.49G>A, Exon 2) BRAF p.V600E (COSM476*, COSV56056643*, substitution, c.1799T>A, Exon 15)

ERBB2 p.E770 A771insAYVM (new: p.Y772 A775dup) (COSM20959*, COSV54062409*,

insertion, c.2313_2324dup, Exon 19)

KRAS p.G12D (COSM521*, COSV55497369*, substitution, c.35G>A, Exon 1) KRAS p.Q61K (COSM549*, COSV55502066*, substitution, c.181C>A, Exon 2) KRAS p.A146T (COSM19404*, COSV55501778*, substitution, c.436G>A, Exon 3) PIK3CA p.C420R (COSM757*, COSV55874020* substitution, c.1258T>C, Exon 7)

PIK3CA p.E542K (COSM760*, COSV55873227*, substitution, c.1624G>A, Exon 9) PIK3CA p.E545A (COSM12458*, COSV55873209*, substitution, c.1634A>C, Exon 9) PIK3CA p.E545D (COSM765*, COSV55874040*, substitution, c.1635G>T, Exon 9)

PIK3CA p.E545G (COSM764*, COSV55873220*, substitution, c.1634A>G, Exon 9) PIK3CA p.E545K (COSM763*, COSV55873239* substitution, c.1633G>A, Exon 9)

PIK3CA p.Q546E (COSM6147*, COSV55882350* substitution, c.1636C>G, Exon 9) PIK3CA p.Q546R (COSM12459*, COSV55876869* substitution, c.1637A>G, Exon 9)

PIK3CA p.H1047L (COSM776*, COSV55873401* substitution, c.3140A>T, Exon 20)

PIK3CA p.H1047R (COSM775*, COSV55873195*, substitution, c.3140A>G, Exon 20)

PIK3CA p.H1047Y (COSM774*, COSV55876499* substitution, c.3139C>T, Exon 20) p.G719S (COSM6252*, COSV51767289*, substitution, c.2155G>A,Exon 18)

p.E746 A750delELREA (COSM6225*, COSV51765066*, deletion, c.2236 2250del15, Exon

p.S752_I759delSPKANKEI (COSM6256*, COSV51774879*, deletion, c.2254_2277del24, Exon 19)

Phone: +49 (0) 381 377 182 01

Net: www.sens-id.com SensID GmbH, Schillingallee 68, 18057 Rostock, Germany Mail: support@sens-id.com



Page 2/4

p.S768I (COSM6241*, COSV51768106* substitution, c.2303G>T, Exon 20)

p.V769_D770insASV (new: p.A767_V769dup) (COSM20884*, COSV51850427* Insertion,

c.2303_2304insTGTGGCCAG, Exon 20)

p.T790M (COSM6240*, COSV51765492*, substitution, c.2369C>T, Exon 20) p.L858R (COSM6224*, COSV51765161*, substitution, c.2573T>G, Exon 21) p.L861Q (COSM6213*, COSV51766344*, substitution, c.2582T>A, Exon 21)

* GRCh38 COSMIC v91

ALLELE FREQUENCY **QUALITY**

0%

DNA quantity metrologically traceable to internationally certified

reference material¹

The copy number values are metrologically traceable to the natural units count 1 and ratio 1 and International System of Units

(SI) derived units of volume.

STORAGE CONDITIONS

+ 2-8 °C

MANUFACTURING AND

SensID GmbH

QUALITY CONTROL

Schillingallee 68, 18057 Rostock, Germany

SITES

TEST METHOD AND	Quality Control	Test Method	Acceptance
ACCEPTANCE CRITERIA			criteria
	Fragmentation	Fragment Length Analysis	peak size 167 bp ±
		Agilent High Sensitivity DNA Kit	10%
		(Agilent Technologies)	(151 bp - 181 bp)
	Quantification	Total DNA measurement:	Total DNA:
		Spectrophotometry	n.a. ⁴
		ssDNA [ng/ μ l] = (A260-A320)*38 ² , ³	
		dsDNA measurement: Qubit dsDNA BR Assay Kit (Invitrogen)	dsDNA: 17.5 – 22.5 ng/μl
	Allele Frequency	ddPCR Analysis using BioRad QX200™ System	AF 0.00% (≤0.03%) ⁵
		•	•

⁴ not applicable

¹ ERM_AD442K

² Protocol NK603 – Community Reference Laboratory for GM Food and Feed

³Measured before filling in product tube

⁵ With the exception of PIK3CA E545A where it is ≤0.7% **Phone**: +49 (0) 381 377 182 01



RESULTS OF ANALYSIS				1 486 57 1
RESULTS OF ANALYSIS		Result		PASS/FAIL
	Fragmentation	172 bp		PASS
	Quantity	33.5 ng/μl (total DNA)		
		19.2 ng/µl (dsDNA)		PASS
		Mutation	AF in %	PASS/FAIL
		AKT1 E17K	0.00	PASS
		BRAF V600E	0.00	PASS
		ERBB2 E770_A771insAYVM (Y772_A775dup)	0.00	PASS
		KRAS G12D	0.00	PASS
		KRAS Q61K	0.00	PASS
		KRAS A146T	0.00	PASS
		PIK3CA C420R	0.00	PASS
	Allele Frequency	PIK3CA E542K	0.00	PASS
		PIK3CA E545A ⁶	0.59	PASS
		PIK3CA E545D	0.00	PASS
		PIK3CA E545G	0.00	PASS
		PIK3CA E545K	0.00	PASS
		PIK3CA Q546E	0.00	PASS
		PIK3CA Q546R	0.00	PASS
		PIK3CA H1047L	0.00	PASS
		PIK3CA H1047R	0.00	PASS
		PIK3CA H1047Y	0.00	PASS
		EGFR L858R	0.00	PASS
		EGFR L861Q	0.00	PASS
		EGFR S768I	0.00	PASS
		EGFR E746_A750delELREA	0.00	PASS
		EGFR T790M	0.00	PASS
		EGFR G719S	0.00	PASS
		EGFR V769_D770insASV	0.00	PASS
		EGFR S752_I759delSPANKEI	0.00	PASS

Net: www.sens-id.com SensID GmbH, Schillingallee 68, 18057 Rostock, Germany

⁶A BLAST sequence analysis shows 98% homology of PIK3CA E545A mutation sequence to genome locus Homo sapiens chromosome 22, GRCh38,p13. Therefore, a higher false positive rate is expected and measured, most likely due to a cross reaction of gene probe to genome locus Homo sapiens chromosome 22, GRCh38,p13.

Phone: +49 (0) 381 377 182 01

Net: www.sens-id.com

Mail: support@sens-id.com



COMMENTS/REMARKS

ADDITIONAL INFORMATION:

Copy numbers (CN) of the respective measurements

Mutation	CN wt ⁷ /μl	CN mut ⁸ /μl
AKT1 E17K	1155	0
BRAF V600E	1012	0
ERBB2 E770_A771insAYVM	1763	0
(Y772_A775dup)		
KRAS G12D	1554	0
KRAS Q61K	1759	0
KRAS A146T	2004	0
PIK3CA C420R	1119	0
PIK3CA E542K	2593	0
PIK3CA E545A ⁶	3185	19
PIK3CA E545D	1790	0
PIK3CA E545G	2619	0
PIK3CA E545K	1462	0
PIK3CA Q546E	2961	0
PIK3CA Q546R	3173	0
PIK3CA H1047L	2348	0
PIK3CA H1047R	2276	0
PIK3CA H1047Y	2236	0
EGFR L858R	2053	0
EGFR L861Q	2909	0
EGFR S768I	1600	0
EGFR E746_A750delELREA	2228	0
EGFR T790M	2247	0
EGFR G719S	2634	0
EGFR V769_D770insASV	1831	0
EGFR S752_I759delSPANKEI	1282	O

Table 1 indicates the values of the QC assays performed by SensID GmbH with an DNA input of ~40 ng. The value for the respective mutation results from the mean value of eight measured replicates (CN values are rounded). CN concentration values per microliter (μ I), are based on droplet digital (ddPCR) assay counts dilution factors, and droplet volume measurements. The detection of the amount of CNs may vary depending on the assay used. Therefore, due to assay properties, there may be deviations in the observed number of copies and allele frequencies compared to the values given here.

Name and position/title of Person authorising the batch release:

Mr. Björn Nowack

Date of batch release: 07.05.2020

Signature batch release: Björn Nowack

This document was created electronically and is valid without a signature.

⁷ Wild Type

⁸ Mutation

Phone: +49 (0) 381 377 182 01

Net: www.sens-id.com SensID GmbH, Schillingallee 68, 18057 Rostock, Germany Mail: support@sens-id.com