

## BATCH CERTIFICATE

For Research Use Only

### PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	1X Plasma (human-tech) DNA free, 40 ml Bulk		
DESCRIPTION	Human serum proteins in common plasma concentrations (1-fold concentrated), electrolytes, EDTA.		
CATALOG NUMBER	SID-000150		
BATCH NUMBER	00882		
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> <li>· Manufactured and sealed in class 2 safety cabinet</li> <li>· Manufactured according to DIN EN ISO 13485:2016</li> </ul>		
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none"> <li>· PCR clean (DNA, DNase, RNase free) LoBind tube with screw cap</li> <li>· Polypropylen (PP)</li> </ul>		
DATE OF MANUFACTURE	12.03.2025		
EXPIRY DATE	11.03.2027		
CONCENTRATION	1X		
NOMINAL VOLUME	40 ml		
QUALITY	DNA-free		
STORAGE CONDITIONS	+ 2-8 °C		
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria
	Quality (DNA-free) LOD	Real-Time PCR Investigator Quantiplex Pro Kit (Qiagen)	< 0.5 pg/μl
RESULTS OF ANALYSIS		Result	PASS / FAIL
	Quality	< 0.5 pg/μl	PASS
COMMENTS / REMARKS	Starting materials were tested and found non-reactive for HbsAG and negative for anti-HIV 1 and 2, anti-HCV and HCV RNA.		

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release: 13.03.2025

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.